


















Symbols Glossary

Stockert GmbH Products

Symbol	Title and the designation number of the SDO-developed standard	Title of the symbol and its reference number,	Meaning or explanatory text of the symbol
	IEC/TR 60878 Ed. 3.0 b:2015 Graphical symbols for electrical equipment in medical practice	Type BF applied part 5.9-5333 IEC 60417-5333	To identify a type BF applied part complying with IEC 60601-1
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Caution 5.4.4	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Date of manufacture 5.1.3	Indicates the date when the medical device was manufactured.
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Manufacturer 5.1.1	Indicates the medical device manufacturer, as defined in EU Directives 90/385/EEC, 93/42/EEC and 98/79/EC.
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Consult instructions for use 5.4.3	Indicates the need for the user to consult the instructions for use.
	IEC/TR 60878 Ed. 3.0 b:2015 Graphical symbols for electrical equipment in medical practice	Stand-by 5.1-5009 IEC 60417-5009	To identify the switch or switch position by means of which part of the equipment is switched on in order to bring it into the stand-by condition

	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Catalogue number 5.1.6	Indicates the manufacturer's catalogue number so that the medical device can be identified.
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Batch code 5.1.5	Indicates the manufacturer's batch code so that the batch or lot can be identified.
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Serial number 5.1.7	Indicates the manufacturer's serial number so that a specific medical device can be identified
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Humidity limitation 5.3.8	Indicates the range of humidity to which the medical device can be safely exposed
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Atmospheric pressure limitation 5.3.9	Indicates the range of atmospheric pressure to which the medical device can be safely exposed
	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Temperature limit 5.3.7	Indicates the temperature limits to which the medical device can be safely exposed
	IEC/TR 60878 Ed. 3.0 b:2015 Graphical symbols for electrical equipment in medical practice	Refer to instruction manual/booklet 5.8- ISO 7010-M002 ISO 7010-M002	To signify that the instruction manual must be read.

	ISO 15223-1 Second Edition 2012-07-01, medical devices—symbols to be used with medical device labels, labelling, and information to be supplied—Part 1: General requirements	Keep dry 5.3.4	Indicates a medical device that needs to be protected from moisture
	IEC/TR 60878 Ed. 3.0 b:2015 Graphical symbols for electrical equipment in medical practice	Variability, rotational adjustment 5.1-1364 ISO 7000-1364	To indicate variability of a quantity. To identify the control by means of which a quantity is increased or decreased.
	IEC/TR 60878 Ed. 3.0 b:2015 Graphical symbols for electrical equipment in medical practice	Fragile; handle with care 5.5-0621 ISO 7000-0621	On transport packaging. To indicate that the content of the package is fragile and that the package must be handled with care.
	IEC/TR 60878 Ed. 3.0 b:2015 Graphical symbols for electrical equipment in medical practice	Remote control, switch on; remote control, activate 5.1-1108 ISO 7000-1108	To identify the control for changing over to remote control